1S480.SVL : Header information	
Device type	SVAN 971
Serial No.	84012
Internal software version	1.14.1
Filesystem version	1,14
Original file name	1S480
Measurement time	02:02:46
Measurement date [d-M-yyyy]	27-11-2022
Device function	1/3 Octave
Leq/RMS integration	Linear
Start delay	0.0s
Integration period	Infinity
Start synchronization	Off
Pre Calibration type	By measurement
Pre Calibration date	26-11-2022
Pre Calibration time	23:02:56
Post Calibration type	By measurement
Post Calibration date	27-11-2022
Post Calibration time	02:46:26
Channel input	Mic.
Measurement range	Low
Mic. field correction	Free
Mic. outdoor filter	Off
Windscreen	On
Pre Calibration factor	1.6 dB
Pre Calibration level	94.0 dB
Post Calibration factor	1.7 dB
Post Calibration level	94.0 dB
Calibration drift	-0.0 dB

Kalibratie informatie

1S480.SVL : Logger results



Bovenruimte: inside blocks is meetdata> 91,9dB

1S480.SVL : Logger results



Kelder ruimte: werkbare meetdata betreft inside blocks 95,6dB