1S478.SVL : Logger results



	Start	Duration	LAeq (TH) [dB]	LCeq (TH) [dB]
Info	-	-	P1 (A, Lin)	P2 (C, Lin)
Main cursor	26-11-2022 23:11:30.800	-	85.4	88.4
Inside blocks	26-11-2022 23:10:39.700	00:01:16.800	82.4	86.0
Outside blocks	26-11-2022 23:10:34.000	00:00:21.200	77.0	80.8
Events	26-11-2022 23:10:34.000	00:01:38.000	-	-

Op meting is muziek overheersend te horen, daarbij stemgeluid.

Zuivere meetdate betreft inside blocks: 82,4 dB

1S478.SVL : Header information	
Device type	SVAN 971
Serial No.	84012
Internal software version	1.14.1
Filesystem version	1,14
Original file name	1S478
Measurement time	23:10:34
Measurement date [d-M-yyyy]	26-11-2022
Device function	1/3 Octave
Leq/RMS integration	Linear
Start delay	0.0s
Integration period	Infinity
Start synchronization	Off
Pre Calibration type	By measurement
Pre Calibration date	26-11-2022
Pre Calibration time	23:02:56
Post Calibration type	By measurement
Post Calibration date	27-11-2022
Post Calibration time	02:46:26
Events	
Event Trigger mode	Whole measure
Sampling rate	12000
Bits per sample	16
Signal Gain	40
Pre trigger	On
Event Filter	Z
Events count	1
Channel input	Mic.
Measurement range	Low
Mic. field correction	Free
Mic. outdoor filter	Off
Windscreen	On
Pre Calibration factor	1.6 dB
Pre Calibration level	94.0 dB
Post Calibration factor	1.7 dB
Post Calibration level	94.0 dB
Calibration drift	-0.0 dB

Kalibratie informatie