To: KeyuserSpiroNose[5.1.2e @ggd.amsterdam.nl]
From: Breathomix Support
Sent: Sat 5/22/2021 4:28:19 PM
Subject: RE: Criteria breathhold
Received: Sat 5/22/2021 4:28:24 PM

Beste,

De breathhold moet inderdaad minimaal 4 seconden zijn.

Met vriendelijke groet

5.1.2e
Support Medewerker

Breathomix BV
Bargelaan 200
2333 CW Leiden
The Netherlands
T: 5.1.2e
E: 5.1.2e @breathomix.com
W: www.breathomix.com

From: KeyuserSpiroNose < 5.1.2e @ggd.amsterdam.nl>
Sent: Saturday, 22 May 2021 17:20
To: Breathomix Support < 5.1.2e @breathomix.com>
Subject: Criteria breathhold

[CAUTION!] Please do not click links or open attachments unless you recognize the sender, this email is from an External Sender.

Beste Lezer,

In de resultaat interpretatie handleiding staat dat de breathhold een dalende lijn van minimaal 5 seconden moet weergeven.
Echter geven een aantal Key Users na hun opleidingsdagen aan dat de criteria nu minimaal 4 seconden is ?

Wat is de juiste criteria om te hanteren voor de breathhold.

Met vriendelijke groet,

5.1.2e
Teststraat RAI
GGD Amsterdam

GGD Amsterdam streeft naar optimale dienstverlening en zorgvuldige afhandeling van e-mailverkeer. Als een e-mail niet voor u is bestemd, verzoekt de gemeente u vriendelijk ons van de juiste adressering op de hoogte te stellen en de e-mail te verwijderen zonder de informatie te gebruiken en te delen met anderen. Voor verdere informatie over de rechten op informatie, zie www.ggd.amsterdam.nl/proclaimer